



CERTIFICATE OF ACCREDITATION

The ANSI National Accreditation Board

Hereby attests that

Precision Calibrations

**221 E. Patti Page Blvd.
Claremore, OK 74017**

Fulfills the requirements of

ISO/IEC 17025:2017

and national standard

ANSI/NCSL Z540-1-1994 (R2002)

In the field of

CALIBRATION

This certificate is valid only when accompanied by a current scope of accreditation document.
The current scope of accreditation can be verified at www.anab.org.

Jason Stine, Vice President

Expiry Date: 28 October 2026

Certificate Number: AC-2092



This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017.
This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory
quality management system (refer to joint ISO-ILAC-IAF Communiqué dated April 2017).

SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017

AND

ANSI/NCSL Z540-1-1994 (R2002)

Precision Calibrations

221 E. Patti Page Blvd
Claremore, OK 74017
Nathan Saylor 918-530-9687

CALIBRATION

Valid to: **October 28, 2026**

Certificate Number: **AC-2092**

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Calibration of Process Instruments (Thermocouple Simulation) ¹	Type J (0 to 1 600) °F	0.8 °F	Thermocouple Simulation using Process Calibrator per AMS2750
	Type K (0 to 2 400) °F	0.88 °F	

Length – Dimensional Metrology

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-) ²	Reference Standard, Method, and/or Equipment
OD Micrometer ¹	(0 to 12) in	(58 + 3.3L) μin	Gage Blocks
Calipers ¹	(0 to 12) in	(570 + 14L) μin	Gage Blocks
Dial Indicator ¹	(0 to 1) in (1 to 4) in	30 μin 580 μin	Gage Blocks
Length Standards	(0 to 12) in	(45 + 3L) μin	Gage Blocks w/indicator

Mass and Mass Related

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Pressure Devices ¹	(0 to 30 000) psi	11 psi	Fluke RPM4
Scales & Balances ¹	(0 to 220) g (0 to 4) kg	1.5 mg 820 mg	OIML Class F2 weights
	(0 to 200) lb	0.1 lb	NIST 105 Class F
Torque Wrenches	(4 to 50) lbf·in (30 to 400) lbf·in (80 to 1 000) lbf·in (20 to 250) lbf·ft	0.9 lbf·in 1.7 lbf·in 4.4 lbf·in 1.8 lbf·ft	Torque Tester and Transducer

Thermodynamic

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
System Accuracy Tests (SAT) ¹	Type J (100 to 1 400) °F	1.7 °F	Process Calibrator w/Thermocouple probes/wire Per AMS2750
	Type K (100 to 2 000) °F	1.8 °F	
Temperature Uniformity Surveys (TUS) ¹	Type J (100 to 500) °F	2.1 °F	Multipoint data logger w/Thermocouple wire per AMS2750
	Type K (200 to 2 000) °F	2.2 °F	
Relative Humidity ¹	(20 to 80) % RH	1.8 %RH	Vaisala with Probe
Temperature ¹	(32 to 572) °F	0.2 °F	Fluke 1552A with probe


Time and Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Stopwatch and Timers ¹	3 600 s	0.35 s	Stopwatch

Calibration and Measurement Capability (CMC) is expressed in terms of the measurement parameter, measurement range, expanded uncertainty of measurement and reference standard, method, and/or equipment. The expanded uncertainty of measurement is expressed as the standard uncertainty of the measurement multiplied by a coverage factor of 2 ($k=2$), corresponding to a confidence level of approximately 95%.

Notes:

1. On-site calibration service is available for this parameter, since on-site conditions are typically more variable than those in the laboratory, larger measurement uncertainties are expected on-site than what is reported on the accredited scope.
2. L = Length in inches.
3. Unless otherwise specified in the far-right column, all calibration procedures utilized during calibration have been written internally.
4. This scope is formatted as part of a single document including Certificate of Accreditation No. AC-2092.



Jason Stine, Vice President

